

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121223\
 Data File : VD077840.D
 Acq On : 12 Dec 2023 14:24
 Operator : JC/SY
 Sample : 05720-02RE
 Misc : 7.20G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TT-FRACSEDIMENT-IDWSO-20231207RE

Quant Time: Dec 12 14:41:48 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 08 13:33:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	136854	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	265825	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	222992	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	72575	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	75556	46.853	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.700%
35) Dibromofluoromethane	7.810	113	84063	45.146	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	90.300%
50) Toluene-d8	10.275	98	334008	44.788	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.580%
62) 4-Bromofluorobenzene	12.575	95	74506	33.416	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	66.840%
Target Compounds						Qvalue
44) Trichloroethene	9.034	130	3792	1.721	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121223\
Data File : VD077840.D
Acq On : 12 Dec 2023 14:24
Operator : JC/SY
Sample : 05720-02RE
Misc : 7.20G/5.0ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TT-FRACSEDIMENT-IDWSO-20231207RE

Quant Time: Dec 12 14:41:48 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 08 13:33:18 2023
Response via : Initial Calibration

