

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121223\
 Data File : VD077841.D
 Acq On : 12 Dec 2023 14:50
 Operator : JC/SY
 Sample : 05692-01RE
 Misc : 5.10G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-03-12072023

Quant Time: Dec 12 15:07:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 08 13:33:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	194170	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	364301	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	340846	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	147371	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	94530	41.315	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.640%
35) Dibromofluoromethane	7.810	113	108367	42.467	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.940%
50) Toluene-d8	10.275	98	465585	45.555	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.100%
62) 4-Bromofluorobenzene	12.575	95	127517	41.732	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.460%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121223\
Data File : VD077841.D
Acq On : 12 Dec 2023 14:50
Operator : JC/SY
Sample : 05692-01RE
Misc : 5.10G/5.0ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-03-12072023

Quant Time: Dec 12 15:07:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 08 13:33:18 2023
Response via : Initial Calibration

