


Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121223\
Data File : VD077831.D
Acq On : 12 Dec 2023 08:24
Operator : JC/SY
Sample : BFB
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Title : SW846 8260
Last Update : Fri Dec 08 13:33:18 2023



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.5	13712	PASS
75	95	30	60	54.2	36280	PASS
95	95	100	100	100.0	66944	PASS
96	95	5	9	6.3	4219	PASS
173	174	0.00	2	2.0	1184	PASS
174	95	50	100	90.5	60592	PASS
175	174	5	9	7.4	4477	PASS
176	174	95	101	96.4	58411	PASS
177	176	5	9	6.7	3923	PASS