

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD121318\
 Data File : VD060559.D
 Acq On : 13 Dec 2018 14:21
 Operator : VA/AP
 Sample : J6199-07
 Misc : 6.16g/5ml/MSVOA D/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-04-121218-B

Quant Time: Dec 14 01:39:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 06 01:49:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.39	168	1534453	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	7.43	114	2149053	50.00	ug/l	-0.03
63) Chlorobenzene-d5	11.27	117	1636136	50.00	ug/l	-0.02
72) 1,4-Dichlorobenzene-d4	13.36	152	682275	50.00	ug/l	-0.02
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.77	65	430608	39.59	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	79.18%	
35) Dibromofluoromethane	6.31	113	609684	38.62	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	77.24%	
50) Toluene-d8	9.43	98	1547870	33.89	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	67.78%	
62) 4-Bromofluorobenzene	12.41	95	541167	33.11	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	66.22%	
Target Compounds						
16) Acetone	3.16	43	63543	36.449	ug/l	95
20) Methylene Chloride	3.72	84	207726	9.340	ug/l	97
52) Toluene	9.52	92	109151	3.253	ug/l	96
68) m/p-Xylenes	11.54	106	79842	3.655	ug/l	90
69) o-Xylene	11.91	106	46018	2.211	ug/l	74
80) 1,3,5-Trimethylbenzene	12.78	105	73045	1.980	ug/l	97
84) 1,2,4-Trimethylbenzene	13.08	105	277960	7.277	ug/l	97
95) Naphthalene	14.96	128	58242	3.335	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD121318\
Data File : VD060559.D
Acq On : 13 Dec 2018 14:21
Operator : VA/AP
Sample : J6199-07
Misc : 6.16g/5ml/MSVOA D/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-04-121218-B

Quant Time: Dec 14 01:39:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 06 01:49:20 2018
Response via : Initial Calibration

