

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD121318\
 Data File : VD060564.D
 Acq On : 13 Dec 2018 16:42
 Operator : VA/AP
 Sample : J6373-01
 Misc : 5.51g/5ml/MSVOA D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP-A-FRAME

Quant Time: Dec 14 01:48:34 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 06 01:49:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.38	168	1580185	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	7.42	114	2193799	50.00	ug/l	-0.03
63) Chlorobenzene-d5	11.27	117	1690066	50.00	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.36	152	723387	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.76	65	414253	36.99	ug/l	-0.03
Spiked Amount 50.000			Recovery	=	73.98%	
35) Dibromofluoromethane	6.31	113	593784	36.85	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	73.70%	
50) Toluene-d8	9.44	98	1555114	33.35	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	66.70%	
62) 4-Bromofluorobenzene	12.41	95	552030	33.08	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	66.16%	
Target Compounds						
16) Acetone	3.16	43	70282	39.147	ug/l	97
52) Toluene	9.52	92	296076	8.644	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD121318\
Data File : VD060564.D
Acq On : 13 Dec 2018 16:42
Operator : VA/AP
Sample : J6373-01
Misc : 5.51g/5ml/MSVOA D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP-A-FRAME

Quant Time: Dec 14 01:48:34 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 06 01:49:20 2018
Response via : Initial Calibration

