

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121323\
 Data File : VD077859.D
 Acq On : 13 Dec 2023 13:50
 Operator : RP/MD
 Sample : 05737-01RE
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-02-12082023RE

Quant Time: Dec 13 13:55:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 08 13:33:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	48719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	88910	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	91142	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	42096	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	35900	62.535	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	125.060%
35) Dibromofluoromethane	7.810	113	30767	49.402	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.800%
50) Toluene-d8	10.275	98	114267	45.811	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.620%
62) 4-Bromofluorobenzene	12.575	95	35343	47.393	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121323\
Data File : VD077859.D
Acq On : 13 Dec 2023 13:50
Operator : RP/MD
Sample : 05737-01RE
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-02-12082023RE

Quant Time: Dec 13 13:55:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 08 13:33:18 2023
Response via : Initial Calibration

