

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121323\
 Data File : VD077861.D
 Acq On : 13 Dec 2023 14:44
 Operator : RP/MD
 Sample : 05699-01RE
 Misc : 5.08G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 ELM-N7THSTREET-DRUMS-SOILRE

Quant Time: Dec 14 01:52:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 08 13:33:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	17442	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	30693	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.580	117	32498	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	14799	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	11361	55.277	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 110.560%	
35) Dibromofluoromethane	7.810	113	11607	53.988	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 107.980%	
50) Toluene-d8	10.269	98	37940	44.061	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 88.120%	
62) 4-Bromofluorobenzene	12.575	95	11292	43.863	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 87.720%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121323\
Data File : VD077861.D
Acq On : 13 Dec 2023 14:44
Operator : RP/MD
Sample : 05699-01RE
Misc : 5.08G/5.0ml/MSVOA_D/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
ELM-N7THSTREET-DRUMS-SOILRE

Quant Time: Dec 14 01:52:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 08 13:33:18 2023
Response via : Initial Calibration

