

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121323\
 Data File : VD077864.D
 Acq On : 13 Dec 2023 16:04
 Operator : RP/MD
 Sample : 05760-05
 Misc : 5.05G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SLAB

Quant Time: Dec 14 01:53:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 08 13:33:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	16770	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	29589	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	31622	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	14167	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	12770	64.622	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.240%
35) Dibromofluoromethane	7.804	113	8538	41.195	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	82.380%
50) Toluene-d8	10.269	98	38389	46.246	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	92.500%
62) 4-Bromofluorobenzene	12.575	95	11882	47.876	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	95.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121323\
Data File : VD077864.D
Acq On : 13 Dec 2023 16:04
Operator : RP/MD
Sample : 05760-05
Misc : 5.05G/5.0ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SLAB

Quant Time: Dec 14 01:53:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 08 13:33:18 2023
Response via : Initial Calibration

