

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121323\
 Data File : VD077868.D
 Acq On : 13 Dec 2023 17:52
 Operator : RP/MD
 Sample : 05611-03
 Misc : 5.04G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PE-3

Quant Time: Dec 14 01:55:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 08 13:33:18 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	187654	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	329954	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	312245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	132476	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	78949	35.704	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	71.400%
35) Dibromofluoromethane	7.805	113	97521	42.195	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.380%
50) Toluene-d8	10.269	98	418251	45.183	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.360%
62) 4-Bromofluorobenzene	12.575	95	114263	41.287	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121323\
Data File : VD077868.D
Acq On : 13 Dec 2023 17:52
Operator : RP/MD
Sample : 05611-03
Misc : 5.04G/5.0ml/MSVOA_D/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PE-3

Quant Time: Dec 14 01:55:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 08 13:33:18 2023
Response via : Initial Calibration

