

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121421\
 Data File : VD071314.D
 Acq On : 14 Dec 2021 12:11
 Operator : VA/SY
 Sample : M5022-09RE
 Misc : 5.01G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP10RE

Quant Time: Dec 15 08:27:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	50896	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	86668	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	109707	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	59278	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	43914	73.532	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	147.060%
35) Dibromofluoromethane	7.920	113	32669	59.194	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	118.380%
50) Toluene-d8	10.337	98	118598	56.469	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	112.940%
62) 4-Bromofluorobenzene	12.626	95	55614	76.391	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	152.780%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121421\
Data File : VD071314.D
Acq On : 14 Dec 2021 12:11
Operator : VA/SY
Sample : M5022-09RE
Misc : 5.01G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP10RE

Quant Time: Dec 15 08:27:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

