

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121421\
 Data File : VD071319.D
 Acq On : 14 Dec 2021 14:31
 Operator : VA/SY
 Sample : M5022-59
 Misc : 5.05G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 TP15RE

Quant Time: Dec 15 08:29:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 12/15/2021
 Supervised By :Mahesh Dadoda 12/17/2021

12/15/2021
 Supervised By :Mahesh
 Dadoda

12/17/2021

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	58616	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.861	114	98948	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	123831	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	59786	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	47886	69.623	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163		Recovery =		139.240%
35) Dibromofluoromethane	7.914	113	37342	59.265	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147		Recovery =		118.520%
50) Toluene-d8	10.338	98	133406	55.636	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140		Recovery =		111.280%
62) 4-Bromofluorobenzene	12.620	95	60020	72.211	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144		Recovery =		144.420%#
Target Compounds						
11) Tert butyl alcohol	5.173	59	8282m	164.447	ug/l	Qvalue
16) Acetone	4.162	43	14079	98.389	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121421\
Data File : VD071319.D
Acq On : 14 Dec 2021 14:31
Operator : VA/SY
Sample : M5022-59
Misc : 5.05G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
TP15RE

Quant Time: Dec 15 08:29:37 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 12/15/2021
Supervised By :Mahesh Dadoda 12/17/2021

12/15/2021
Supervised By :Mahesh
Dadoda

12/17/2021

