

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121421\
 Data File : VD071330.D
 Acq On : 14 Dec 2021 19:41
 Operator : VA/SY
 Sample : M5065-01RE
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 UV-1WCRE

Quant Time: Dec 15 08:33:46 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	41786	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	71029	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	90017	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	46469	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	32913	67.127	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.260%
35) Dibromofluoromethane	7.914	113	23360	51.647	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.300%
50) Toluene-d8	10.332	98	91702	53.276	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.560%
62) 4-Bromofluorobenzene	12.620	95	44092	73.899	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	147.800%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121421\
Data File : VD071330.D
Acq On : 14 Dec 2021 19:41
Operator : VA/SY
Sample : M5065-01RE
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
UV-1WCRE

Quant Time: Dec 15 08:33:46 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

