

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121421\
 Data File : VD071310.D
 Acq On : 14 Dec 2021 10:20
 Operator : VA/SY
 Sample : VD1214SBL01
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1214SBL01

Quant Time: Dec 15 08:24:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	250766	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	447242	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	406781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	185250	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	165738	56.327	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	112.660%
35) Dibromofluoromethane	7.914	113	145996	51.263	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.520%
50) Toluene-d8	10.338	98	500747	46.203	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	92.400%
62) 4-Bromofluorobenzene	12.626	95	176369	46.946	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.900%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121421\
Data File : VD071310.D
Acq On : 14 Dec 2021 10:20
Operator : VA/SY
Sample : VD1214SBL01
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1214SBL01

Quant Time: Dec 15 08:24:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration





Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.979 min Scan# 1030

Delta R.T. -0.006 min

Lab File: VD071310.D

Acq: 14 Dec 2021 10:20

Instrument :

MSVOA_D

ClientSampleId :





VD1214SBL01

Tgt Ion:168 Resp: 250766

Ion Ratio Lower Upper

168 100

99 63.9 46.6 69.8



1,2-Dichloroethane-d4

Concen: 56.327 ug/l

RT: 8.326 min Scan# 1089

Delta R.T. 0.000 min

Lab File: VD071310.D




Acq: 14 Dec 2021 10:20


Tgt Ion: 65 Resp: 165738

Ion Ratio Lower Upper

65 100

67 51.1 0.0 112.2









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.861 min Scan# 1180
Delta R.T. 0.000 min
Lab File: VD071310.D
Acq: 14 Dec 2021 10:20




Instrument :
MSVOA_D
ClientSampleId :
VD1214SBL01


Tgt Ion:114 Resp: 447242
Ion Ratio Lower Upper
114 100
63 23.6 0.0 41.8
88 15.9 0.0 32.8



#35
Dibromofluoromethane
Concen: 51.263 ug/l
RT: 7.914 min Scan# 1019
Delta R.T. -0.006 min
Lab File: VD071310.D
Acq: 14 Dec 2021 10:20

Tgt Ion:113 Resp: 145996
Ion Ratio Lower Upper
113 100
111 106.2 82.0 123.0
192 17.8 15.2 22.8









#50
Toluene-d8
Concen: 46.203 ug/l
RT: 10.338 min Scan# 1431
Delta R.T. -0.000 min
Lab File: VD071310.D
Acq: 14 Dec 2021 10:20




Instrument :
MSVOA_D
ClientSampleId :
VD1214SBL01


Tgt Ion: 98 Resp: 500747
Ion Ratio Lower Upper
98 100
100 64.9 51.8 77.6



#62
4-Bromofluorobenzene
Concen: 46.946 ug/l
RT: 12.626 min Scan# 1820
Delta R.T. -0.000 min
Lab File: VD071310.D
Acq: 14 Dec 2021 10:20

Tgt Ion: 95 Resp: 176369
Ion Ratio Lower Upper
95 100
174 77.3 0.0 159.4
176 72.4 0.0 154.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.638 min Scan# 117
Delta R.T. -0.000 min
Lab File: VD071310.D
Acq: 14 Dec 2021 10:20

Instrument :
MSVOA_D
ClientSampleId :
VD1214SBL01

Tgt Ion:117 Resp: 406781
Ion Ratio Lower Upper
117 100
82 61.3 44.7 67.1
119 32.9 26.2 39.4



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.567 min Scan# 1980
Delta R.T. 0.000 min
Lab File: VD071310.D
Acq: 14 Dec 2021 10:20

Tgt Ion:152 Resp: 185250
Ion Ratio Lower Upper
152 100
115 61.1 29.8 89.3
150 159.7 0.0 349.2

