

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121423\
 Data File : VD077884.D
 Acq On : 14 Dec 2023 15:36
 Operator : RP/MD
 Sample : 05788-12
 Misc : 5.40G/10ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 DCLP0

Quant Time: Dec 15 07:13:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM112923SMA.M
 Quant Title : SFAM01.0
 QLast Update : Wed Dec 06 00:24:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	8.781	114	327551	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	307166	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.516	152	130080	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.281	65	218470	24.458	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	97.840%
7) Chloroethane-d5	2.811	69	201384	28.285	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	113.160%
11) 1,1-Dichloroethene-d2	3.928	65	55263	23.861	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	95.440%
21) 2-Butanone-d5	6.993	46	45542	40.341	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.680%
24) Chloroform-d	7.575	84	259985	29.450	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	117.800%
26) 1,2-Dichloroethane-d4	8.234	65	122981	27.495	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	109.960%
32) Benzene-d6	8.205	84	559792	30.260	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	121.040%
36) 1,2-Dichloropropane-d6	9.210	67	165680	29.352	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	117.400%
41) Toluene-d8	10.269	98	515224	30.258	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	121.040%
43) trans-1,3-Dichloroprop...	10.528	79	64095	25.679	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	102.720%
47) 2-Hexanone-d5	10.875	63	39053	40.151	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	80.300%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	97085	23.802	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	95.200%
66) 1,2-Dichlorobenzene-d4	13.816	152	139778	30.037	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	120.160%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121423\
Data File : VD077884.D
Acq On : 14 Dec 2023 15:36
Operator : RP/MD
Sample : 05788-12
Misc : 5.40G/10ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
DCLP0

Quant Time: Dec 15 07:13:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM112923SMA.M
Quant Title : SFAM01.0
QLast Update : Wed Dec 06 00:24:24 2023
Response via : Initial Calibration

