

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121521\
 Data File : VD071337.D
 Acq On : 15 Dec 2021 12:40
 Operator : VA/SY
 Sample : M5073-01RE
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EO-02-121321RE

Quant Time: Dec 16 04:26:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	15456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	26440	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	28449	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	15484	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	12186	67.193	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	134.380%
35) Dibromofluoromethane	7.908	113	8076	47.967	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.940%
50) Toluene-d8	10.332	98	32503	50.729	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	101.460%
62) 4-Bromofluorobenzene	12.626	95	13621	61.329	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	122.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121521\
Data File : VD071337.D
Acq On : 15 Dec 2021 12:40
Operator : VA/SY
Sample : M5073-01RE
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EO-02-121321RE

Quant Time: Dec 16 04:26:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

