

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121521\
 Data File : VD071339.D
 Acq On : 15 Dec 2021 13:36
 Operator : VA/SY
 Sample : M5064-01RE
 Misc : 5.33G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SASP2-121-1-(0-5)RE

Quant Time: Dec 16 04:27:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	31380	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.861	114	50665	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	64319	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	32362	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.332	65	26094	70.868	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	141.740%
35) Dibromofluoromethane	7.920	113	18377	56.960	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	113.920%
50) Toluene-d8	10.332	98	64133	52.235	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.480%
62) 4-Bromofluorobenzene	12.626	95	30452	71.552	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	143.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121521\
Data File : VD071339.D
Acq On : 15 Dec 2021 13:36
Operator : VA/SY
Sample : M5064-01RE
Misc : 5.33G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SASP2-121-1-(0-5)RE

Quant Time: Dec 16 04:27:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

