

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121521\
 Data File : VD071355.D
 Acq On : 15 Dec 2021 21:06
 Operator : VA/SY
 Sample : M5105-01
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HD-01-121521

Quant Time: Dec 16 04:33:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	302576	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	519271	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	599211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	315158	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	152052	42.827	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	85.660%
35) Dibromofluoromethane	7.914	113	157920	47.758	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	95.520%
50) Toluene-d8	10.332	98	678692	53.935	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.860%
62) 4-Bromofluorobenzene	12.620	95	296530	67.981	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	135.960%
Target Compounds						
95) Naphthalene	15.379	128	38726	4.209	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121521\
Data File : VD071355.D
Acq On : 15 Dec 2021 21:06
Operator : VA/SY
Sample : M5105-01
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HD-01-121521

Quant Time: Dec 16 04:33:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

