



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121522\
Data File : VD075022.D
Acq On : 15 Dec 2022 11:36
Operator : KP/SY
Sample : BFB
Misc : 5.00G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D121522S.M
Title : SW846 8260
Last Update : Fri Dec 16 05:30:45 2022



AutoFind: Scans 1867, 1868, 1869; Background Corrected with Scan 1859

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.8	10029	PASS
75	95	30	60	48.4	30733	PASS
95	95	100	100	100.0	63469	PASS
96	95	5	9	6.8	4291	PASS
173	174	0.00	2	1.4	821	PASS
174	95	50	100	89.6	56840	PASS
175	174	5	9	7.5	4263	PASS
176	174	95	101	98.7	56075	PASS
177	176	5	9	6.9	3855	PASS