

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121523\
 Data File : VD077903.D
 Acq On : 15 Dec 2023 15:02
 Operator : RP/MD
 Sample : 05833-01
 Misc : 5.02G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 351

Quant Time: Dec 16 04:02:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	9102	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	15824	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	15023	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	6217	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	5693	53.080	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	106.160%
35) Dibromofluoromethane	7.805	113	5263	47.482	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.960%
50) Toluene-d8	10.269	98	19208	43.267	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	86.540%
62) 4-Bromofluorobenzene	12.581	95	6286	47.361	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	94.720%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121523\
Data File : VD077903.D
Acq On : 15 Dec 2023 15:02
Operator : RP/MD
Sample : 05833-01
Misc : 5.02G/5.0ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
351

Quant Time: Dec 16 04:02:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

