

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121523\
 Data File : VD077906.D
 Acq On : 15 Dec 2023 16:23
 Operator : RP/MD
 Sample : 05624-01
 Misc : 5.04G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 14TH-SOIL-1

Quant Time: Dec 16 04:03:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	200628	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	356269	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	332652	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	146952	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	90424	38.249	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	76.500%
35) Dibromofluoromethane	7.810	113	105881	42.428	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	84.860%
50) Toluene-d8	10.269	98	451036	45.126	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.260%
62) 4-Bromofluorobenzene	12.575	95	127095	42.532	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121523\
Data File : VD077906.D
Acq On : 15 Dec 2023 16:23
Operator : RP/MD
Sample : 05624-01
Misc : 5.04G/5.0ml/MSVOA_D/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
14TH-SOIL-1

Quant Time: Dec 16 04:03:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

