

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121523\
 Data File : VD077924.D
 Acq On : 16 Dec 2023 02:44
 Operator : RP/MD
 Sample : 05765-02
 Misc : 5.03G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RT-SB-02-10-11

Quant Time: Dec 16 05:13:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	165739	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.775	114	312787	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	299611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	131277	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	88910	45.525	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.040%
35) Dibromofluoromethane	7.810	113	95452	43.566	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.140%
50) Toluene-d8	10.269	98	392600	44.740	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.480%
62) 4-Bromofluorobenzene	12.575	95	113452	43.244	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.480%
Target Compounds						
17) Carbon Disulfide	4.275	76	15097	2.240	ug/l	# 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121523\
Data File : VD077924.D
Acq On : 16 Dec 2023 02:44
Operator : RP/MD
Sample : 05765-02
Misc : 5.03G/5.0ml/MSVOA_D/SOIL
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RT-SB-02-10-11

Quant Time: Dec 16 05:13:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

