

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD121619\
 Data File : VD064526.D
 Acq On : 16 Dec 2019 16:32
 Operator : VA/SY
 Sample : K6325-04
 Misc : 7.41G/5.00ml/MSVOA D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RT-3425

Quant Time: Dec 17 03:32:18 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 29 07:18:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.98	168	23921	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.87	114	36901	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.65	117	40381	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.58	152	24140	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.33	65	15809	73.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	146.02%	
35) Dibromofluoromethane	7.92	113	12514	53.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.64%	
50) Toluene-d8	10.34	98	45995	52.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.04%	
62) 4-Bromofluorobenzene	12.64	95	19746	70.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	140.32%	
Target Compounds						
67) Ethyl Benzene	11.74	91	20468	13.659	ug/l	95
68) m/p-Xylenes	11.85	106	14613	25.576	ug/l	93
69) o-Xylene	12.18	106	7637	14.473	ug/l	93
70) Styrene	12.20	104	6347	6.922	ug/l #	82
80) 1,3,5-Trimethylbenzene	12.97	105	41199	29.038	ug/l	95
84) 1,2,4-Trimethylbenzene	13.27	105	92176	64.611	ug/l	100
95) Naphthalene	15.41	128	13689750	15401.717	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD121619\
Data File : VD064526.D
Acq On : 16 Dec 2019 16:32
Operator : VA/SY
Sample : K6325-04
Misc : 7.41G/5.00ml/MSVOA D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RT-3425

Quant Time: Dec 17 03:32:18 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D112719S.M
Quant Title : SW846 8260
QLast Update : Fri Nov 29 07:18:03 2019
Response via : Initial Calibration

