

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121621\
 Data File : VD071370.D
 Acq On : 16 Dec 2021 16:36
 Operator : VA/SY
 Sample : M5111-03
 Misc : 5.95G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-3

Quant Time: Dec 16 23:50:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	169808	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	304991	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	295645	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	135327	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	148749	74.654	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	149.300%
35) Dibromofluoromethane	7.909	113	111435	57.377	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	114.760%
50) Toluene-d8	10.332	98	393609	53.256	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.520%
62) 4-Bromofluorobenzene	12.620	95	136841	53.413	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.820%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121621\
Data File : VD071370.D
Acq On : 16 Dec 2021 16:36
Operator : VA/SY
Sample : M5111-03
Misc : 5.95G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-3

Quant Time: Dec 16 23:50:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

