

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121621\
 Data File : VD071372.D
 Acq On : 16 Dec 2021 17:32
 Operator : VA/SY
 Sample : M5124-01
 Misc : 5.03G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 F2337-STOPJOINT-TP

Quant Time: Dec 16 23:51:08 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	375061	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	657114	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.637	117	580686	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	250986	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	239588	54.441	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.880%	
35) Dibromofluoromethane	7.908	113	210976	50.419	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.840%	
50) Toluene-d8	10.332	98	836699	52.544	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 105.080%	
62) 4-Bromofluorobenzene	12.620	95	267078	48.385	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 96.780%	
Target Compounds						Qvalue
16) Acetone	4.156	43	8511	37.182	ug/l	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121621\
Data File : VD071372.D
Acq On : 16 Dec 2021 17:32
Operator : VA/SY
Sample : M5124-01
Misc : 5.03G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
F2337-STOPJOINT-TP

Quant Time: Dec 16 23:51:08 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

