

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121623\
 Data File : VD077929.D
 Acq On : 16 Dec 2023 12:50
 Operator : RP/MD
 Sample : VD1216SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1216SBL01

Quant Time: Dec 18 01:56:26 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	190713	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.775	114	340446	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	324659	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	147617	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	89293	39.734	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	79.460%
35) Dibromofluoromethane	7.805	113	101814	42.695	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.380%
50) Toluene-d8	10.269	98	433639	45.402	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.800%
62) 4-Bromofluorobenzene	12.575	95	123771	43.344	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121623\
Data File : VD077929.D
Acq On : 16 Dec 2023 12:50
Operator : RP/MD
Sample : VD1216SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1216SBL01

Quant Time: Dec 18 01:56:26 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

