

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121623\
 Data File : VD077939.D
 Acq On : 16 Dec 2023 17:40
 Operator : RP/MD
 Sample : 05853-05
 Misc : 4.10G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EAST-WALL

Quant Time: Dec 18 02:01:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.875	168	177101	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.781	114	316763	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	297916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	125898	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	80570	38.608	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	77.220%
35) Dibromofluoromethane	7.810	113	94561	42.618	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.240%
50) Toluene-d8	10.269	98	404334	45.499	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.000%
62) 4-Bromofluorobenzene	12.575	95	110762	41.689	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121623\
Data File : VD077939.D
Acq On : 16 Dec 2023 17:40
Operator : RP/MD
Sample : 05853-05
Misc : 4.10G/5.0ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EAST-WALL

Quant Time: Dec 18 02:01:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

