

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121623\
 Data File : VD077941.D
 Acq On : 16 Dec 2023 18:34
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 18 02:02:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	0.625	0.463	25.9#	58	0.00
3 P	Chloromethane	0.851	0.851	0.0	77	0.00
4 C	Vinyl Chloride	1.156	1.174	-1.6#	75	0.00
5 T	Bromomethane	0.649	0.768	-18.3	86	0.00
6 T	Chloroethane	0.788	0.839	-6.5	79	0.00
7 T	Trichlorofluoromethane	1.064	0.912	14.3	64	0.00
8 T	Diethyl Ether	0.299	0.295	1.3	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.609	0.528	13.3	64	0.00
10 T	Methyl Iodide	0.399	0.586	-46.9#	82	0.00
11 T	Tert butyl alcohol	0.032	0.030	6.3	83	0.00
12 CM	1,1-Dichloroethene	0.601	0.518	13.8#	63	0.00
13 T	Acrolein	0.062	0.048	22.6	59	0.00
14 T	Allyl chloride	0.827	0.708	14.4	64	0.00
15 T	Acrylonitrile	0.128	0.133	-3.9	78	0.00
16 T	Acetone	0.161	0.124	23.0	79	0.00
17 T	Carbon Disulfide	2.033	1.697	16.5	62	0.00
18 T	Methyl Acetate	0.464	0.559	-20.5	91	0.00
19 T	Methyl tert-butyl Ether	1.275	1.320	-3.5	77	0.00
20 T	Methylene Chloride	0.835	0.739	11.5	77	0.00
21 T	trans-1,2-Dichloroethene	0.681	0.644	5.4	67	0.00
22 T	Diisopropyl ether	1.727	1.728	-0.1	72	0.00
23 T	Vinyl Acetate	0.844	0.793	6.0	68	0.00
24 P	1,1-Dichloroethane	1.170	1.160	0.9	73	0.00
25 T	2-Butanone	0.185	0.158	14.6	78	0.00
26 T	2,2-Dichloropropane	0.982	0.880	10.4	66	0.00
27 T	cis-1,2-Dichloroethene	0.733	0.748	-2.0	74	0.00
28 T	Bromochloromethane	0.473	0.506	-7.0	79	0.00
29 T	Tetrahydrofuran	0.094	0.097	-3.2	78	0.00
30 C	Chloroform	1.171	1.198	-2.3#	75	0.00
31 T	Cyclohexane	1.048	0.850	18.9	61	0.00
32 T	1,1,1-Trichloroethane	1.032	0.954	7.6	67	0.00
33 S	1,2-Dichloroethane-d4	0.589	0.603	-2.4	75	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	85	0.00
35 S	Dibromofluoromethane	0.350	0.347	0.9	78	0.00
36 T	1,1-Dichloropropene	0.560	0.470	16.1	65	0.00
37 T	Ethyl Acetate	0.218	0.188	13.8	77	0.00
38 T	Carbon Tetrachloride	0.558	0.471	15.6	65	0.00
39 T	Methylcyclohexane	0.664	0.558	16.0	63	0.00
40 TM	Benzene	1.590	1.485	6.6	73	0.00
41 T	Methacrylonitrile	0.111	0.111	0.0	87	0.00
42 TM	1,2-Dichloroethane	0.412	0.398	3.4	76	0.00
43 T	Isopropyl Acetate	0.388	0.374	3.6	79	0.00
44 TM	Trichloroethene	0.414	0.373	9.9	69	0.00
45 C	1,2-Dichloropropane	0.387	0.376	2.8#	76	0.00
46 T	Dibromomethane	0.205	0.209	-2.0	80	0.00
47 T	Bromodichloromethane	0.518	0.524	-1.2	78	0.00
48 T	Methyl methacrylate	0.221	0.195	11.8	71	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121623\
 Data File : VD077941.D
 Acq On : 16 Dec 2023 18:34
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleID :
 VSTDCCC050

Quant Time: Dec 18 02:02:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	82	0.00
50 S	Toluene-d8	1.403	1.371	2.3	74	0.00
51 T	4-Methyl-2-Pentanone	0.198	0.192	3.0	77	0.00
52 CM	Toluene	0.986	0.937	5.0#	72	0.00
53 T	t-1,3-Dichloropropene	0.498	0.489	1.8	76	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.595	1.2	77	0.00
55 T	1,1,2-Trichloroethane	0.280	0.291	-3.9	82	0.00
56 T	Ethyl methacrylate	0.247	0.254	-2.8	80	0.00
57 T	1,3-Dichloropropane	0.482	0.486	-0.8	79	0.00
58 T	2-Chloroethyl Vinyl ether	0.141	0.141	0.0	86	0.00
59 T	2-Hexanone	0.164	0.140	14.6	78	0.00
60 T	Dibromochloromethane	0.337	0.347	-3.0	80	0.00
61 T	1,2-Dibromoethane	0.263	0.264	-0.4	80	0.00
62 S	4-Bromofluorobenzene	0.419	0.441	-5.3	82	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	90	0.00
64 T	Tetrachloroethene	0.367	0.313	14.7	70	0.00
65 PM	Chlorobenzene	1.310	1.033	21.1	75	0.00
66 T	1,1,1,2-Tetrachloroethane	0.416	0.384	7.7	77	0.00
67 C	Ethyl Benzene	2.082	1.858	10.8#	73	0.00
68 T	m/p-Xylenes	0.796	0.718	9.8	73	0.00
69 T	o-Xylene	0.734	0.681	7.2	76	0.00
70 T	Styrene	1.091	1.044	4.3	77	0.00
71 P	Bromoform	0.217	0.212	2.3	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
73 T	Isopropylbenzene	3.790	3.348	11.7	72	0.00
74 T	N-amyl acetate	0.745	0.627	15.8	74	0.00
75 P	1,1,2,2-Tetrachloroethane	0.661	0.641	3.0	82	0.00
76 T	1,2,3-Trichloropropane	0.434	0.414	4.6	71	0.00
77 T	Bromobenzene	0.891	0.834	6.4	80	0.00
78 T	n-propylbenzene	4.618	4.145	10.2	73	0.00
79 T	2-Chlorotoluene	2.475	2.246	9.3	76	0.00
80 T	1,3,5-Trimethylbenzene	3.161	2.827	10.6	73	0.00
81 T	trans-1,4-Dichloro-2-butene	0.214	0.200	6.5	80	0.00
82 T	4-Chlorotoluene	2.587	2.377	8.1	76	0.00
83 T	tert-Butylbenzene	2.668	2.493	6.6	75	0.00
84 T	1,2,4-Trimethylbenzene	3.161	2.947	6.8	76	0.00
85 T	sec-Butylbenzene	3.944	3.586	9.1	73	0.00
86 T	p-Isopropyltoluene	3.531	3.224	8.7	73	0.00
87 T	1,3-Dichlorobenzene	1.823	1.712	6.1	78	0.00
88 T	1,4-Dichlorobenzene	1.966	1.681	14.5	77	0.00
89 T	n-Butylbenzene	3.323	3.058	8.0	73	0.00
90 T	Hexachloroethane	0.649	0.632	2.6	80	0.00
91 T	1,2-Dichlorobenzene	1.550	1.471	5.1	81	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.102	0.094	7.8	83	0.00
93 T	1,2,4-Trichlorobenzene	1.059	0.968	8.6	80	0.00
94 T	Hexachlorobutadiene	0.571	0.492	13.8	71	0.00
95 T	Naphthalene	1.759	1.918	-9.0	92	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121623\
Data File : VD077941.D
Acq On : 16 Dec 2023 18:34
Operator : RP/MD
Sample : VSTDCCC050
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Dec 18 02:02:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.873	0.826	5.4	82	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6