

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD121718\
 Data File : VD060603.D
 Acq On : 17 Dec 2018 16:59
 Operator : VA/AP
 Sample : J6424-03
 Misc : 4.81g/5ml/MSVOA D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EB09_1.5-2.5

Quant Time: Dec 18 13:23:29 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 06 01:49:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.42	168	1176531	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	7.47	114	1716617	50.00	ug/l	0.01
63) Chlorobenzene-d5	11.29	117	1453376	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.38	152	739142	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.81	65	507678	60.88	ug/l	0.01
Spiked Amount 50.000			Recovery	=	121.76%	
35) Dibromofluoromethane	6.34	113	647482	51.35	ug/l	0.01
Spiked Amount 50.000			Recovery	=	102.70%	
50) Toluene-d8	9.47	98	1808394	49.56	ug/l	0.01
Spiked Amount 50.000			Recovery	=	99.12%	
62) 4-Bromofluorobenzene	12.44	95	654795	50.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.30%	
Target Compounds						
8) Diethyl Ether	2.85	74	5357	2.784	ug/l	78
16) Acetone	3.18	43	42183	31.557	ug/l #	91
20) Methylene Chloride	3.75	84	243960	16.917	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD121718\
Data File : VD060603.D
Acq On : 17 Dec 2018 16:59
Operator : VA/AP
Sample : J6424-03
Misc : 4.81g/5ml/MSVOA D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EB09_1.5-2.5

Quant Time: Dec 18 13:23:29 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 06 01:49:20 2018
Response via : Initial Calibration

