

Data Path : Z:\VOASRV\HPCHEM1\MSVOA D\DATA\VD121718\
 Data File : VD060611.D
 Acq On : 17 Dec 2018 20:45
 Operator : VA/AP
 Sample : J6445-06
 Misc : 4.6µl/5ml/MSVOA D/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 EB05_1.5-3.5

Quant Time: Dec 18 13:44:42 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 06 01:49:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.42	168	849362	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	7.46	114	1282504	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.30	117	1011872	50.00	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.38	152	439201	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.80	65	439378	72.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	145.96%	
35) Dibromofluoromethane	6.34	113	542725	57.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	115.22%	
50) Toluene-d8	9.46	98	1369105	50.22	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.44%	
62) 4-Bromofluorobenzene	12.44	95	434724	44.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.14%	
Target Compounds						
16) Acetone	3.17	43	16749	17.357	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA D\DATA\VD121718\
Data File : VD060611.D
Acq On : 17 Dec 2018 20:45
Operator : VA/AP
Sample : J6445-06
Misc : 4.6µl/5ml/MSVOA D/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
EB05_1.5-3.5

Quant Time: Dec 18 13:44:42 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 06 01:49:20 2018
Response via : Initial Calibration

