


Data Path : Z:\voasrv\HPCHEM1\MSVOA D\Data\VD121718\
Data File : VD060590.D
Acq On : 17 Dec 2018 10:20
Operator : VA/AP
Sample : BFB
Misc : 5.00µl/5ml/MSVOA D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
Title : SW846 8260
Last Update : Thu Dec 06 01:49:20 2018



AutoFind: Scans 1108, 1109, 1110; Background Corrected with Scan 1103

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.1	94936	PASS
75	95	30	60	41.3	194930	PASS
95	95	100	100	100.0	471957	PASS
96	95	5	9	6.5	30759	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	89.2	421154	PASS
175	174	5	9	7.2	30277	PASS
176	174	95	101	99.2	417586	PASS
177	176	5	9	6.3	26117	PASS