

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121721\
 Data File : VD071380.D
 Acq On : 17 Dec 2021 12:52
 Operator : VA/SY
 Sample : M5116-02
 Misc : 5.06G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VOC

Quant Time: Dec 19 01:52:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	404608	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.856	114	697184	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	610514	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	254930	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.320	65	245015	51.608	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 103.220%	
35) Dibromofluoromethane	7.909	113	219405	49.420	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 98.840%	
50) Toluene-d8	10.332	98	865512	51.229	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 102.460%	
62) 4-Bromofluorobenzene	12.620	95	260917	44.552	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.100%	
Target Compounds						Qvalue
20) Methylene Chloride	4.968	84	39477	4.434	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121721\
Data File : VD071380.D
Acq On : 17 Dec 2021 12:52
Operator : VA/SY
Sample : M5116-02
Misc : 5.06G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VOC

Quant Time: Dec 19 01:52:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

