

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121721\
 Data File : VD071382.D
 Acq On : 17 Dec 2021 14:54
 Operator : VA/SY
 Sample : M5130-01RE
 Misc : 6.86G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BLUE-GREEN-CLAYRE

Quant Time: Dec 19 01:53:04 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	310869	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.856	114	567630	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	370426	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	56110	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	211359	57.943	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.880%
35) Dibromofluoromethane	7.909	113	190743	52.770	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.540%
50) Toluene-d8	10.332	98	611931	44.486	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	88.980%
62) 4-Bromofluorobenzene	12.620	95	123153	25.828	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	51.660%
Target Compounds						
					Qvalue	
16) Acetone	4.150	43	58081	83.374	ug/l	92
25) 2-Butanone	7.214	43	10314	36.533	ug/l	92
44) Trichloroethene	9.103	130	4752	1.237	ug/l	76
64) Tetrachloroethene	10.873	164	19527	8.304	ug/l	96
84) 1,2,4-Trimethylbenzene	13.255	105	9349	2.858	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121721\
Data File : VD071382.D
Acq On : 17 Dec 2021 14:54
Operator : VA/SY
Sample : M5130-01RE
Misc : 6.86G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BLUE-GREEN-CLAYRE

Quant Time: Dec 19 01:53:04 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

