

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121721\
 Data File : VD071383.D
 Acq On : 17 Dec 2021 16:12
 Operator : VA/SY
 Sample : M5146-01
 Misc : 5.04G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 NB-8-121721

Quant Time: Dec 19 01:53:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.967	168	344044	50.000	ug/l	-0.02
34) 1,4-Difluorobenzene	8.855	114	606891	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	556128	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	240193	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	232309	57.546	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.100%
35) Dibromofluoromethane	7.908	113	200410	51.858	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.720%
50) Toluene-d8	10.332	98	769132	52.298	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.600%
62) 4-Bromofluorobenzene	12.620	95	241713	47.414	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	94.820%
Target Compounds						
16) Acetone	4.156	43	6506	36.118	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121721\
Data File : VD071383.D
Acq On : 17 Dec 2021 16:12
Operator : VA/SY
Sample : M5146-01
Misc : 5.04G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
NB-8-121721

Quant Time: Dec 19 01:53:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

