

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121721\
 Data File : VD071384.D
 Acq On : 17 Dec 2021 16:41
 Operator : VA/SY
 Sample : M5147-01
 Misc : 5.02G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VNJ-210

Quant Time: Dec 19 01:53:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.973	168	380404	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.861	114	651985	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	582244	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	248338	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	242351	54.295	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.580%	
35) Dibromofluoromethane	7.908	113	216819	52.223	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 104.440%	
50) Toluene-d8	10.332	98	828158	52.416	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 104.840%	
62) 4-Bromofluorobenzene	12.620	95	254476	46.465	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 92.920%	
Target Compounds						
16) Acetone	4.156	43	4292	33.971	ug/l	Qvalue # 40

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121721\
Data File : VD071384.D
Acq On : 17 Dec 2021 16:41
Operator : VA/SY
Sample : M5147-01
Misc : 5.02G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VNJ-210

Quant Time: Dec 19 01:53:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

