



Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121721\
 Data File : VD071375.D
 Acq On : 17 Dec 2021 08:57
 Operator : VA/SY
 Sample : BFB
 Misc : 5.00G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Title : SW846 8260
 Last Update : Mon Nov 29 17:54:09 2021



AutoFind: Scans 1818, 1819, 1820; Background Corrected with Scan 1810

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	23.4	59261	PASS
75	95	30	60	56.4	143109	PASS
95	95	100	100	100.0	253525	PASS
96	95	5	9	6.4	16298	PASS
173	174	0.00	2	2.0	3692	PASS
174	95	50	100	72.9	184704	PASS
175	174	5	9	8.4	15550	PASS
176	174	95	101	95.7	176768	PASS
177	176	5	9	7.2	12690	PASS