

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121823\
 Data File : VD077950.D
 Acq On : 18 Dec 2023 16:32
 Operator : RP/MD
 Sample : 05808-11
 Misc : 5.23G/10ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 C0B05

Quant Time: Dec 18 23:18:29 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM112923SMA.M
 Quant Title : SFAM01.0
 QLast Update : Mon Dec 18 23:16:23 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	8.775	114	327959	25.000	ug/L	0.00
28) Chlorobenzene-d5	11.581	117	274319	25.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	13.522	152	74007	25.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	2.275	65	191004	21.357	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	85.440%
7) Chloroethane-d5	2.805	69	180612	25.336	ug/L	0.00
Spiked Amount	25.000	Range	30 - 150	Recovery	=	101.360%
11) 1,1-Dichloroethene-d2	3.922	65	39110	16.865	ug/L	0.00
Spiked Amount	25.000	Range	45 - 110	Recovery	=	67.480%
21) 2-Butanone-d5	6.981	46	44601	39.458	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	78.920%
24) Chloroform-d	7.569	84	203960	23.075	ug/L	0.00
Spiked Amount	25.000	Range	40 - 150	Recovery	=	92.280%
26) 1,2-Dichloroethane-d4	8.228	65	101542	22.673	ug/L	0.00
Spiked Amount	25.000	Range	70 - 130	Recovery	=	90.680%
32) Benzene-d6	8.199	84	405895	24.568	ug/L	0.00
Spiked Amount	25.000	Range	20 - 135	Recovery	=	98.280%
36) 1,2-Dichloropropane-d6	9.210	67	125969	24.989	ug/L	0.00
Spiked Amount	25.000	Range	70 - 120	Recovery	=	99.960%
41) Toluene-d8	10.269	98	335941	22.091	ug/L	0.00
Spiked Amount	25.000	Range	30 - 130	Recovery	=	88.360%
43) trans-1,3-Dichloroprop...	10.522	79	49062	22.009	ug/L	0.00
Spiked Amount	25.000	Range	30 - 135	Recovery	=	88.040%
47) 2-Hexanone-d5	10.881	63	42038	48.395	ug/L	0.00
Spiked Amount	50.000	Range	20 - 135	Recovery	=	96.800%
56) 1,1,2,2-Tetrachloroeth...	12.651	84	84694	23.250	ug/L	0.00
Spiked Amount	25.000	Range	45 - 120	Recovery	=	93.000%
66) 1,2-Dichlorobenzene-d4	13.816	152	60743	22.943	ug/L	0.00
Spiked Amount	25.000	Range	75 - 120	Recovery	=	91.760%
Target Compounds						
13) Acetone	4.022	43	10803	7.459	ug/L	62
65) 1,4-Dichlorobenzene	13.545	146	2051	0.414	ug/L	88
67) 1,2-Dichlorobenzene	13.834	146	2980	0.693	ug/L	89
70) 1,2,4-trichlorobenzene	15.098	180	6951	2.517	ug/L #	68
72) 1,2,3-Trichlorobenzene	15.522	180	3119	1.288	ug/L #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121823\
Data File : VD077950.D
Acq On : 18 Dec 2023 16:32
Operator : RP/MD
Sample : 05808-11
Misc : 5.23G/10ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
C0B05

Quant Time: Dec 18 23:18:29 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\SFAMDLM112923SMA.M
Quant Title : SFAM01.0
QLast Update : Mon Dec 18 23:16:23 2023
Response via : Initial Calibration

