

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD121918\
 Data File : VD060640.D
 Acq On : 19 Dec 2018 15:07
 Operator : VA/AP
 Sample : J6194-07
 Misc : 19.46G/5ml/MSVOA_D/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 OK-01-121818-A

Quant Time: Dec 20 01:27:17 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 06 01:49:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.41	168	20028	50.00	ug/l	-0.01
34) 1,4-Difluorobenzene	7.44	114	33376	50.00	ug/l	-0.01
63) Chlorobenzene-d5	11.28	117	32088	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.37	152	14725	50.00	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.78	65	13680	96.37	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	192.74%	
35) Dibromofluoromethane	6.32	113	11827	48.24	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	96.48%	
50) Toluene-d8	9.45	98	23212	32.72	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	65.44%	
62) 4-Bromofluorobenzene	12.42	95	10495	41.34	ug/l	-0.01
Spiked Amount 50.000			Recovery	=	82.68%	
Target Compounds						
20) Methylene Chloride	3.72	84	3655	14.300	ug/l #	81
52) Toluene	9.55	92	3998	7.672	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD121918\
Data File : VD060640.D
Acq On : 19 Dec 2018 15:07
Operator : VA/AP
Sample : J6194-07
Misc : 19.46G/5ml/MSVOA_D/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
OK-01-121818-A

Quant Time: Dec 20 01:27:17 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 06 01:49:20 2018
Response via : Initial Calibration

