

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD121918\
 Data File : VD060642.D
 Acq On : 19 Dec 2018 16:04
 Operator : VA/AP
 Sample : J6199-11
 Misc : 17.66G/5ml/MSVOA_D/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SU-02-121818-B

Quant Time: Dec 20 13:39:31 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 06 01:49:20 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.39	168	18049	50.00	ug/l	-0.03
34) 1,4-Difluorobenzene	7.44	114	27728	50.00	ug/l	-0.02
63) Chlorobenzene-d5	11.27	117	25665	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.37	152	12341	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.78	65	10924	85.39	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	170.78%	
35) Dibromofluoromethane	6.31	113	11394	55.94	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	111.88%	
50) Toluene-d8	9.44	98	21178	35.93	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	71.86%	
62) 4-Bromofluorobenzene	12.42	95	7689	36.46	ug/l	-0.02
Spiked Amount 50.000			Recovery	=	72.92%	
Target Compounds						
16) Acetone	3.15	43	8198	399.781	ug/l	Qvalue 96
20) Methylene Chloride	3.72	84	3033	12.779	ug/l #	80
80) 1,3,5-Trimethylbenzene	12.79	105	1546	2.317	ug/l	100
84) 1,2,4-Trimethylbenzene	13.10	105	5796	8.388	ug/l	88
89) n-Butylbenzene	13.65	91	1895	2.011	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD121918\
Data File : VD060642.D
Acq On : 19 Dec 2018 16:04
Operator : VA/AP
Sample : J6199-11
Misc : 17.66G/5ml/MSVOA_D/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SU-02-121818-B

Quant Time: Dec 20 13:39:31 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D120518S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 06 01:49:20 2018
Response via : Initial Calibration

