

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121923\
 Data File : VD077964.D
 Acq On : 19 Dec 2023 10:25
 Operator : RP/MD
 Sample : VD1219SBL01
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1219SBL01

Quant Time: Dec 20 01:59:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	249947	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	436804	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.587	117	416285	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	188063	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	116452	39.539	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	79.080%
35) Dibromofluoromethane	7.810	113	135358	44.240	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.480%
50) Toluene-d8	10.275	98	551179	44.978	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.960%
62) 4-Bromofluorobenzene	12.575	95	163089	44.514	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121923\
Data File : VD077964.D
Acq On : 19 Dec 2023 10:25
Operator : RP/MD
Sample : VD1219SBL01
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1219SBL01

Quant Time: Dec 20 01:59:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

