

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121923\
 Data File : VD077981.D
 Acq On : 19 Dec 2023 18:27
 Operator : RP/MD
 Sample : 05765-03
 Misc : 5.06G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RT-SB-03-9.5-10.5

Quant Time: Dec 20 02:08:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	162698	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	298858	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	257210	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	82204	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	75685	39.478	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	78.960%
35) Dibromofluoromethane	7.810	113	87261	41.684	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	83.360%
50) Toluene-d8	10.269	98	368191	43.914	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	87.820%
62) 4-Bromofluorobenzene	12.575	95	87584	34.940	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	69.880%
Target Compounds						
17) Carbon Disulfide	4.281	76	11522	1.741	ug/l	Qvalue # 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121923\
Data File : VD077981.D
Acq On : 19 Dec 2023 18:27
Operator : RP/MD
Sample : 05765-03
Misc : 5.06G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RT-SB-03-9.5-10.5

Quant Time: Dec 20 02:08:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

