

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121923\
 Data File : VD077963.D
 Acq On : 19 Dec 2023 09:52
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 20 01:58:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	85	0.00
2 T	Dichlorodifluoromethane	0.625	0.501	19.8	70	0.00
3 P	Chloromethane	0.851	0.796	6.5	80	0.00
4 C	Vinyl Chloride	1.156	1.173	-1.5#	84	0.00
5 T	Bromomethane	0.649	0.904	-39.3#	114	0.00
6 T	Chloroethane	0.788	0.856	-8.6	90	0.00
7 T	Trichlorofluoromethane	1.064	1.043	2.0	81	0.00
8 T	Diethyl Ether	0.299	0.301	-0.7	83	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.609	0.647	-6.2	87	0.01
10 T	Methyl Iodide	0.399	0.566	-41.9#	88	0.00
11 T	Tert butyl alcohol	0.032	0.030	6.3	90	0.00
12 CM	1,1-Dichloroethene	0.601	0.584	2.8#	80	0.00
13 T	Acrolein	0.062	0.037	40.3#	51	0.00
14 T	Allyl chloride	0.827	0.755	8.7	76	0.00
15 T	Acrylonitrile	0.128	0.132	-3.1	87	0.00
16 T	Acetone	0.161	0.150	6.8	107	0.00
17 T	Carbon Disulfide	2.033	1.725	15.2	70	0.00
18 T	Methyl Acetate	0.464	0.521	-12.3	94	0.00
19 T	Methyl tert-butyl Ether	1.275	1.364	-7.0	89	0.00
20 T	Methylene Chloride	0.835	0.715	14.4	83	0.00
21 T	trans-1,2-Dichloroethene	0.681	0.711	-4.4	83	0.01
22 T	Diisopropyl ether	1.727	1.729	-0.1	81	0.00
23 T	Vinyl Acetate	0.844	0.796	5.7	77	0.00
24 P	1,1-Dichloroethane	1.170	1.198	-2.4	84	0.00
25 T	2-Butanone	0.185	0.179	3.2	99	0.00
26 T	2,2-Dichloropropane	0.982	1.032	-5.1	87	0.00
27 T	cis-1,2-Dichloroethene	0.733	0.806	-10.0	89	0.00
28 T	Bromochloromethane	0.473	0.492	-4.0	86	0.00
29 T	Tetrahydrofuran	0.094	0.093	1.1	83	0.00
30 C	Chloroform	1.171	1.272	-8.6#	89	0.00
31 T	Cyclohexane	1.048	0.948	9.5	76	0.00
32 T	1,1,1-Trichloroethane	1.032	1.111	-7.7	87	0.00
33 S	1,2-Dichloroethane-d4	0.589	0.565	4.1	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.350	0.363	-3.7	84	0.00
36 T	1,1-Dichloropropene	0.560	0.578	-3.2	82	0.00
37 T	Ethyl Acetate	0.218	0.182	16.5	77	0.00
38 T	Carbon Tetrachloride	0.558	0.598	-7.2	86	0.00
39 T	Methylcyclohexane	0.664	0.684	-3.0	80	0.00
40 TM	Benzene	1.590	1.681	-5.7	85	0.00
41 T	Methacrylonitrile	0.111	0.121	-9.0	98	0.00
42 TM	1,2-Dichloroethane	0.412	0.444	-7.8	88	0.00
43 T	Isopropyl Acetate	0.388	0.388	0.0	84	0.00
44 TM	Trichloroethene	0.414	0.438	-5.8	84	0.00
45 C	1,2-Dichloropropane	0.387	0.431	-11.4#	90	0.00
46 T	Dibromomethane	0.205	0.240	-17.1	95	0.00
47 T	Bromodichloromethane	0.518	0.593	-14.5	92	0.00
48 T	Methyl methacrylate	0.221	0.220	0.5	82	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121923\
 Data File : VD077963.D
 Acq On : 19 Dec 2023 09:52
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 20 01:58:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	107	0.00
50 S	Toluene-d8	1.403	1.437	-2.4	80	0.00
51 T	4-Methyl-2-Pentanone	0.198	0.203	-2.5	84	0.00
52 CM	Toluene	0.986	1.086	-10.1#	86	0.00
53 T	t-1,3-Dichloropropene	0.498	0.553	-11.0	89	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.661	-9.8	88	0.00
55 T	1,1,2-Trichloroethane	0.280	0.322	-15.0	94	0.00
56 T	Ethyl methacrylate	0.247	0.270	-9.3	88	0.00
57 T	1,3-Dichloropropane	0.482	0.535	-11.0	89	0.00
58 T	2-Chloroethyl Vinyl ether	0.141	0.103	27.0#	65	0.00
59 T	2-Hexanone	0.164	0.158	3.7	90	0.00
60 T	Dibromochloromethane	0.337	0.406	-20.5	97	0.00
61 T	1,2-Dibromoethane	0.263	0.300	-14.1	94	0.00
62 S	4-Bromofluorobenzene	0.419	0.437	-4.3	84	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	87	0.00
64 T	Tetrachloroethene	0.367	0.391	-6.5	86	0.00
65 PM	Chlorobenzene	1.310	1.259	3.9	89	0.00
66 T	1,1,1,2-Tetrachloroethane	0.416	0.471	-13.2	92	0.00
67 C	Ethyl Benzene	2.082	2.326	-11.7#	90	0.00
68 T	m/p-Xylenes	0.796	0.902	-13.3	89	0.00
69 T	o-Xylene	0.734	0.815	-11.0	88	0.00
70 T	Styrene	1.091	1.275	-16.9	92	0.00
71 P	Bromoform	0.217	0.255	-17.5	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	93	0.00
73 T	Isopropylbenzene	3.790	3.927	-3.6	89	0.00
74 T	N-amyl acetate	0.745	0.662	11.1	82	0.00
75 P	1,1,2,2-Tetrachloroethane	0.661	0.685	-3.6	92	0.00
76 T	1,2,3-Trichloropropane	0.434	0.512	-18.0	92	0.00
77 T	Bromobenzene	0.891	0.917	-2.9	92	0.00
78 T	n-propylbenzene	4.618	4.778	-3.5	88	0.00
79 T	2-Chlorotoluene	2.475	2.547	-2.9	90	0.00
80 T	1,3,5-Trimethylbenzene	3.161	3.329	-5.3	90	0.00
81 T	trans-1,4-Dichloro-2-butene	0.214	0.212	0.9	90	0.00
82 T	4-Chlorotoluene	2.587	2.731	-5.6	92	0.00
83 T	tert-Butylbenzene	2.668	2.851	-6.9	91	0.00
84 T	1,2,4-Trimethylbenzene	3.161	3.378	-6.9	91	0.00
85 T	sec-Butylbenzene	3.944	4.305	-9.2	93	0.00
86 T	p-Isopropyltoluene	3.531	3.889	-10.1	92	0.00
87 T	1,3-Dichlorobenzene	1.823	1.984	-8.8	94	0.00
88 T	1,4-Dichlorobenzene	1.966	1.943	1.2	94	0.00
89 T	n-Butylbenzene	3.323	3.595	-8.2	90	0.00
90 T	Hexachloroethane	0.649	0.747	-15.1	99	0.00
91 T	1,2-Dichlorobenzene	1.550	1.677	-8.2	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.102	0.101	1.0	94	0.00
93 T	1,2,4-Trichlorobenzene	1.059	1.051	0.8	92	0.00
94 T	Hexachlorobutadiene	0.571	0.603	-5.6	91	0.00
95 T	Naphthalene	1.759	1.871	-6.4	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD121923\
Data File : VD077963.D
Acq On : 19 Dec 2023 09:52
Operator : RP/MD
Sample : VSTDCCC050
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Dec 20 01:58:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.873	0.900	-3.1	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6