

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD122018\
 Data File : VD060666.D
 Acq On : 20 Dec 2018 18:52
 Operator : VA/AP
 Sample : J6481-04
 Misc : 5.03g/5ml/MSVOA_D/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 HOUSE-WALL

Quant Time: Dec 21 06:03:28 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122018S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 21 03:34:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.41	168	1415133	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	7.45	114	1895411	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.29	117	1373749	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.37	152	527588	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.79	65	390782	39.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	79.92%	
35) Dibromofluoromethane	6.33	113	544743	39.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	78.54%	
50) Toluene-d8	9.45	98	1361089	32.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	65.76%	
62) 4-Bromofluorobenzene	12.43	95	431234	29.75	ug/l	0.00
Spiked Amount 50.000			Recovery	=	59.50%	
Target Compounds						
20) Methylene Chloride	3.74	84	52038	2.934	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD122018\
Data File : VD060666.D
Acq On : 20 Dec 2018 18:52
Operator : VA/AP
Sample : J6481-04
Misc : 5.03g/5ml/MSVOA_D/SOIL
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
HOUSE-WALL

Quant Time: Dec 21 06:03:28 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122018S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 21 03:34:39 2018
Response via : Initial Calibration

