

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122023\
 Data File : VD077990.D
 Acq On : 20 Dec 2023 12:47
 Operator : RP/MD
 Sample : BLK
 Misc : 5.04G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 BLK

Quant Time: Dec 21 05:22:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	185051	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	333851	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	325600	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	149893	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	86302	39.578	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	79.160%
35) Dibromofluoromethane	7.810	113	101882	43.567	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	87.140%
50) Toluene-d8	10.275	98	420172	44.861	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	89.720%
62) 4-Bromofluorobenzene	12.575	95	125560	44.839	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122023\
Data File : VD077990.D
Acq On : 20 Dec 2023 12:47
Operator : RP/MD
Sample : BLK
Misc : 5.04G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
BLK

Quant Time: Dec 21 05:22:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

