

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122023\
 Data File : VD077997.D
 Acq On : 20 Dec 2023 15:56
 Operator : RP/MD
 Sample : 05827-10RE
 Misc : 5.06G/5.0ml/MSVOA_D/SOIL/B
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 RT-SB-10-11-12RE

Quant Time: Dec 21 05:25:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	162450	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	302310	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	283417	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	117776	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	79068	41.305	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	82.620%
35) Dibromofluoromethane	7.810	113	80554	38.041	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	76.080%
50) Toluene-d8	10.269	98	385935	45.505	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.000%
62) 4-Bromofluorobenzene	12.575	95	108559	42.813	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.620%
Target Compounds						
17) Carbon Disulfide	4.275	76	9759	1.477	ug/l	Qvalue # 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122023\
Data File : VD077997.D
Acq On : 20 Dec 2023 15:56
Operator : RP/MD
Sample : 05827-10RE
Misc : 5.06G/5.0ml/MSVOA_D/SOIL/B
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
RT-SB-10-11-12RE

Quant Time: Dec 21 05:25:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

