

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122023\
 Data File : VD078007.D
 Acq On : 20 Dec 2023 21:44
 Operator : RP/MD
 Sample : VD1220SBL02
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 VD1220SBL02

Quant Time: Dec 21 05:30:23 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	178259	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	324094	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	306557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.516	152	137733	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.228	65	86039	40.961	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.920%
35) Dibromofluoromethane	7.810	113	96655	42.576	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.160%
50) Toluene-d8	10.269	98	410831	45.184	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.360%
62) 4-Bromofluorobenzene	12.575	95	118816	43.709	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	87.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122023\
Data File : VD078007.D
Acq On : 20 Dec 2023 21:44
Operator : RP/MD
Sample : VD1220SBL02
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
VD1220SBL02

Quant Time: Dec 21 05:30:23 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

