

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122023\
 Data File : VD078014.D
 Acq On : 21 Dec 2023 01:46
 Operator : RP/MD
 Sample : 05876-04
 Misc : 4.13G/5.0ml/MSVOA_D/SOIL/A
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 COLW8

Quant Time: Dec 21 05:33:56 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.881	168	167017	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.781	114	302575	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.581	117	285988	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.522	152	118481	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.234	65	79152	40.218	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	80.440%
35) Dibromofluoromethane	7.810	113	91123	42.994	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	85.980%
50) Toluene-d8	10.269	98	384901	45.343	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.680%
62) 4-Bromofluorobenzene	12.575	95	107922	42.524	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.040%
Target Compounds						
95) Naphthalene	15.328	128	28236	6.773	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122023\
Data File : VD078014.D
Acq On : 21 Dec 2023 01:46
Operator : RP/MD
Sample : 05876-04
Misc : 4.13G/5.0ml/MSVOA_D/SOIL/A
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
COLW8

Quant Time: Dec 21 05:33:56 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

