

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122023\
 Data File : VD078004.D
 Acq On : 20 Dec 2023 19:03
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 05:28:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	70	0.00
2 T	Dichlorodifluoromethane	0.625	0.460	26.4#	53	0.00
3 P	Chloromethane	0.851	0.791	7.1	66	0.00
4 C	Vinyl Chloride	1.156	1.146	0.9#	67	0.00
5 T	Bromomethane	0.649	0.866	-33.4#	90	0.00
6 T	Chloroethane	0.788	0.866	-9.9	75	0.00
7 T	Trichlorofluoromethane	1.064	0.959	9.9	62	0.00
8 T	Diethyl Ether	0.299	0.304	-1.7	69	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.609	0.607	0.3	67	0.01
10 T	Methyl Iodide	0.399	0.651	-63.2#	84	0.00
11 T	Tert butyl alcohol	0.032	0.028	12.5	70	0.00
12 CM	1,1-Dichloroethene	0.601	0.567	5.7#	64	0.00
13 T	Acrolein	0.062	0.037	40.3#	42#	0.00
14 T	Allyl chloride	0.827	0.751	9.2	62	0.00
15 T	Acrylonitrile	0.128	0.137	-7.0	74	0.00
16 T	Acetone	0.161	0.112	30.4#	66	0.00
17 T	Carbon Disulfide	2.033	1.656	18.5	56	0.00
18 T	Methyl Acetate	0.464	0.535	-15.3	80	0.00
19 T	Methyl tert-butyl Ether	1.275	1.358	-6.5	73	0.00
20 T	Methylene Chloride	0.835	0.719	13.9	69	0.00
21 T	trans-1,2-Dichloroethene	0.681	0.687	-0.9	66	0.00
22 T	Diisopropyl ether	1.727	1.759	-1.9	68	0.00
23 T	Vinyl Acetate	0.844	0.807	4.4	64	0.00
24 P	1,1-Dichloroethane	1.170	1.211	-3.5	70	0.00
25 T	2-Butanone	0.185	0.157	15.1	71	0.00
26 T	2,2-Dichloropropane	0.982	0.933	5.0	65	0.00
27 T	cis-1,2-Dichloroethene	0.733	0.802	-9.4	73	0.00
28 T	Bromochloromethane	0.473	0.550	-16.3	79	0.00
29 T	Tetrahydrofuran	0.094	0.097	-3.2	72	0.00
30 C	Chloroform	1.171	1.287	-9.9#	74	0.00
31 T	Cyclohexane	1.048	0.881	15.9	58	0.00
32 T	1,1,1-Trichloroethane	1.032	1.023	0.9	66	0.00
33 S	1,2-Dichloroethane-d4	0.589	0.669	-13.6	77	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	76	0.00
35 S	Dibromofluoromethane	0.350	0.416	-18.9	83	0.00
36 T	1,1-Dichloropropene	0.560	0.527	5.9	65	0.00
37 T	Ethyl Acetate	0.218	0.203	6.9	74	0.00
38 T	Carbon Tetrachloride	0.558	0.530	5.0	66	0.00
39 T	Methylcyclohexane	0.664	0.590	11.1	60	0.00
40 TM	Benzene	1.590	1.587	0.2	70	0.00
41 T	Methacrylonitrile	0.111	0.113	-1.8	79	0.00
42 TM	1,2-Dichloroethane	0.412	0.431	-4.6	74	0.00
43 T	Isopropyl Acetate	0.388	0.384	1.0	72	0.00
44 TM	Trichloroethene	0.414	0.409	1.2	68	0.00
45 C	1,2-Dichloropropane	0.387	0.410	-5.9#	74	0.00
46 T	Dibromomethane	0.205	0.220	-7.3	75	0.00
47 T	Bromodichloromethane	0.518	0.569	-9.8	76	0.00
48 T	Methyl methacrylate	0.221	0.223	-0.9	72	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122023\
 Data File : VD078004.D
 Acq On : 20 Dec 2023 19:03
 Operator : RP/MD
 Sample : VSTDCCC050
 Misc : 5.00G/5.0ml/MSVOA_D/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_D
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 21 05:28:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Dec 14 14:48:10 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	85	0.00
50 S	Toluene-d8	1.403	1.573	-12.1	76	0.00
51 T	4-Methyl-2-Pentanone	0.198	0.199	-0.5	71	0.00
52 CM	Toluene	0.986	1.023	-3.8#	70	0.00
53 T	t-1,3-Dichloropropene	0.498	0.514	-3.2	72	0.00
54 T	cis-1,3-Dichloropropene	0.602	0.630	-4.7	73	0.00
55 T	1,1,2-Trichloroethane	0.280	0.318	-13.6	80	0.00
56 T	Ethyl methacrylate	0.247	0.266	-7.7	75	0.00
57 T	1,3-Dichloropropane	0.482	0.519	-7.7	75	0.00
58 T	2-Chloroethyl Vinyl ether	0.141	0.149	-5.7	81	0.00
59 T	2-Hexanone	0.164	0.137	16.5	68	0.00
60 T	Dibromochloromethane	0.337	0.390	-15.7	81	0.00
61 T	1,2-Dibromoethane	0.263	0.295	-12.2	80	0.00
62 S	4-Bromofluorobenzene	0.419	0.479	-14.3	80	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	76	0.00
64 T	Tetrachloroethene	0.367	0.350	4.6	67	0.00
65 PM	Chlorobenzene	1.310	1.155	11.8	71	0.00
66 T	1,1,1,2-Tetrachloroethane	0.416	0.433	-4.1	74	0.00
67 C	Ethyl Benzene	2.082	2.105	-1.1#	71	0.00
68 T	m/p-Xylenes	0.796	0.832	-4.5	72	0.00
69 T	o-Xylene	0.734	0.773	-5.3	73	0.00
70 T	Styrene	1.091	1.182	-8.3	75	0.00
71 P	Bromoform	0.217	0.240	-10.6	80	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
73 T	Isopropylbenzene	3.790	3.572	5.8	70	0.00
74 T	N-amyl acetate	0.745	0.626	16.0	66	0.00
75 P	1,1,2,2-Tetrachloroethane	0.661	0.669	-1.2	77	0.00
76 T	1,2,3-Trichloropropane	0.434	0.365	15.9	56	0.00
77 T	Bromobenzene	0.891	0.858	3.7	74	0.00
78 T	n-propylbenzene	4.618	4.430	4.1	70	0.00
79 T	2-Chlorotoluene	2.475	2.356	4.8	72	0.00
80 T	1,3,5-Trimethylbenzene	3.161	3.042	3.8	71	0.00
81 T	trans-1,4-Dichloro-2-butene	0.214	0.209	2.3	76	0.00
82 T	4-Chlorotoluene	2.587	2.558	1.1	74	0.00
83 T	tert-Butylbenzene	2.668	2.643	0.9	72	0.00
84 T	1,2,4-Trimethylbenzene	3.161	3.133	0.9	73	0.00
85 T	sec-Butylbenzene	3.944	3.890	1.4	72	0.00
86 T	p-Isopropyltoluene	3.531	3.472	1.7	71	0.00
87 T	1,3-Dichlorobenzene	1.823	1.795	1.5	73	0.00
88 T	1,4-Dichlorobenzene	1.966	1.758	10.6	73	0.00
89 T	n-Butylbenzene	3.323	3.183	4.2	69	0.00
90 T	Hexachloroethane	0.649	0.684	-5.4	78	0.00
91 T	1,2-Dichlorobenzene	1.550	1.541	0.6	77	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.102	0.097	4.9	77	0.00
93 T	1,2,4-Trichlorobenzene	1.059	0.930	12.2	70	0.00
94 T	Hexachlorobutadiene	0.571	0.513	10.2	67	0.00
95 T	Naphthalene	1.759	1.774	-0.9	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122023\
Data File : VD078004.D
Acq On : 20 Dec 2023 19:03
Operator : RP/MD
Sample : VSTDCCC050
Misc : 5.00G/5.0ml/MSVOA_D/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_D
LabSampleId :
VSTDCCC050

Quant Time: Dec 21 05:28:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D120823S.M
Quant Title : SW846 8260
QLast Update : Thu Dec 14 14:48:10 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.873	0.833	4.6	74	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6