

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_D\DATA\VD122118\
 Data File : VD060680.D
 Acq On : 21 Dec 2018 13:30
 Operator : VA/AP
 Sample : J6481-03RE
 Misc : 5.09g/5ml/MSVOA_D/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 PROP-SW-1RE

Quant Time: Dec 21 15:13:02 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_D\METHOD\82D122018S.M
 Quant Title : SW846 8260
 QLast Update : Fri Dec 21 03:34:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	6.44	168	1052979	50.00	ug/l	0.03
34) 1,4-Difluorobenzene	7.47	114	1397597	50.00	ug/l	0.03
63) Chlorobenzene-d5	11.30	117	1058100	50.00	ug/l	0.02
72) 1,4-Dichlorobenzene-d4	13.39	152	422321	50.00	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.81	65	286744	39.41	ug/l	0.02
Spiked Amount 50.000			Recovery	=	78.82%	
35) Dibromofluoromethane	6.35	113	394722	38.59	ug/l	0.02
Spiked Amount 50.000			Recovery	=	77.18%	
50) Toluene-d8	9.49	98	951580	31.18	ug/l	0.03
Spiked Amount 50.000			Recovery	=	62.36%	
62) 4-Bromofluorobenzene	12.44	95	305144	28.55	ug/l	0.02
Spiked Amount 50.000			Recovery	=	57.10%	
Target Compounds						
20) Methylene Chloride	3.75	84	13774	1.044	ug/l	92
69) o-Xylene	11.95	106	37634	2.711	ug/l	88
78) n-propylbenzene	12.66	91	67007	1.743	ug/l	97
80) 1,3,5-Trimethylbenzene	12.81	105	128913	5.552	ug/l	97
84) 1,2,4-Trimethylbenzene	13.11	105	85215	3.563	ug/l	97
85) sec-Butylbenzene	13.24	105	38156	1.184	ug/l	99
95) Naphthalene	14.98	128	65352	5.911	ug/l #	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_D\DATA\VD122118\
Data File : VD060680.D
Acq On : 21 Dec 2018 13:30
Operator : VA/AP
Sample : J6481-03RE
Misc : 5.09g/5ml/MSVOA_D/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
PROP-SW-1RE

Quant Time: Dec 21 15:13:02 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_D\METHOD\82D122018S.M
Quant Title : SW846 8260
QLast Update : Fri Dec 21 03:34:39 2018
Response via : Initial Calibration

