

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
 Data File : VD071412.D
 Acq On : 21 Dec 2021 15:57
 Operator : VA/SY
 Sample : M5184-02
 Misc : 5.77G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-2

Quant Time: Dec 21 19:08:44 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	451888	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	815348	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	721167	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.567	152	317045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	267755	50.497	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.000%
35) Dibromofluoromethane	7.914	113	262267	50.513	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.020%
50) Toluene-d8	10.338	98	1027271	51.992	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.980%
62) 4-Bromofluorobenzene	12.626	95	334310	48.812	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	97.620%
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.215	96	9450	1.769	ug/l	93
64) Tetrachloroethene	10.873	164	7976	1.742	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
Data File : VD071412.D
Acq On : 21 Dec 2021 15:57
Operator : VA/SY
Sample : M5184-02
Misc : 5.77G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-2

Quant Time: Dec 21 19:08:44 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

