

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
 Data File : VD071413.D
 Acq On : 21 Dec 2021 16:25
 Operator : VA/SY
 Sample : M5184-05
 Misc : 7.15G/5.00ml/MSVOA_D/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_D
 ClientSampleId :
 SB-5

Quant Time: Dec 21 19:09:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 29 17:54:09 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.979	168	65103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.861	114	114998	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.638	117	114179	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.561	152	52159	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.326	65	62027	81.197	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	162.400%
35) Dibromofluoromethane	7.914	113	44253	60.431	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	120.860%
50) Toluene-d8	10.338	98	147745	53.017	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.040%
62) 4-Bromofluorobenzene	12.626	95	51566	53.381	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.760%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_D\Data\VD122121\
Data File : VD071413.D
Acq On : 21 Dec 2021 16:25
Operator : VA/SY
Sample : M5184-05
Misc : 7.15G/5.00ml/MSVOA_D/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_D
ClientSampleId :
SB-5

Quant Time: Dec 21 19:09:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_D\Method\82D112921S.M
Quant Title : SW846 8260
QLast Update : Mon Nov 29 17:54:09 2021
Response via : Initial Calibration

